



Quality Certificate

VWR* NITRILE EXAM GLOVES

This document certifies that the below product meets the following criteria:

North American Catalog No:	82026-426
Lot Number:	234R10J012
Description:	VWR GLOVE NITRILE PF M PK100

Material : Nitril BUTADIENE Rubber(NBR).

QC Testing : Representative production samples are collected and inspected in accordance with current applicable product specifications:

Meet or exceed ASTM D 6319-00a(Water test & Dimension)
Meet or exceed EN 455-1&2
ISO 2859 (Sampling table)

Accuracy: Water tight test : Inspection level=G1; AQL 1.5
Visual Inspection: Inspection level=G1; AQL 1.5(critical defects)
AQL 2.5(major defects)
AQL 4.0(minor defects)
Dimensional: Inspection level=S2; AQI 4.0
Physical property: Inspection level=S2; AQI 4.0
Powder Content: Inspection level=5 pieces;Average \leq 2mg/glove
Moisture Content: Inspection level=5 pieces;Average \leq 1%
Glove Content: Inspection level=5 dispensers:Tolerance 100+/-2 pieces

Sterility: N/A

Non-Pyrogenic: N/A

Rnase/Dnase Free: N/A

Regulatory: FDA 510K,ISO 13485,CMDCAS Quality Systems,GMP Certificate

Signed:

Ken Crossley
Manager
Quality Assurance

Date of Issue: July 26, 2010