



Quality Certificate

VWR* GLOVES

This document certifies that the below product meets the following criteria:

North American Catalog No:	82026-426
Lot Number:	PT234R11P004
Description:	VWR GLOVE NITRILE PF M PK100
Manufacturing Date:	November 2011
Product shelf Life:	October 2016

Material: Nitrile Butadiene Rubber, 100% latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current applicable product specifications.

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)
Meet or exceed EN 455-1&2
ISO 2859 (sampling table)

Accuracy:

Water tight test:	Inspection level = G1; AQL1.5
Visual inspection:	Inspection level = G1; AQL1.5 (critical defects) AQL2.5 (major defects) AQL4.0 (minor defects)
Dimensional:	Inspection level = S2; AQL4.0
Physical Property:	Inspection level = S2; AQL4.0
Powder Content:	Inspection level = 5pieces; Average ≤ 2 mg/glove
Moisture Content:	Inspection level = 5pieces; Average $\leq 1\%$
Glove Count:	Inspection level = 5 dispensers; Tolerance 100+/-2pieces

Sterility: N/A

Non-Pyrogenic: N/A

Rnase/Dnase Free: N/A

Regulatory: FDA510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

Signed:

Ken Crossley
Manager
Quality Assurance

Date of Issue: November 22, 2011