

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance

Catalog Number: 3022-500
Lot Number: 10113046
Product Description: PYREX® Single Metric Scale, 500mL Graduated Cylinder, TC

Component Materials:

Glass Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E-438 (Standard Specification for Glasses in Laboratory Apparatus)
Guard Polyethylene, Low Density (LDPE)

Attributes: Capacity: 500mL ±4.0mL

This product is designed to comply with the requirements of ASTM E 1272 (Specification for Graduated Cylinders) This product is calibrated in accordance with ASTM E 542 (Specification for Calibration of Laboratory Volumetric Apparatus) to assure conformance to capacity specification. This product complies with all criteria for class B volumetric ware per ASTM E 694 (Specification for Volumetric Ware). This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following Characteristics:

Test/Procedure:	Result:
Visual Attributes	Pass
Dimensional Specification	Pass
Packaging	Pass
Capacity	Pass



Arthur D. Birdsall
Quality Assurance Manager

Rev No: 83706