



## Quality Certificate

### VWR\* GLOVES

This document certifies that the below product meets the following criteria:

<b>North American Catalog No:</b>	82026-426
<b>Lot Number:</b>	PT234R12N105
<b>Description:</b>	VWR GLOVE NITRILE PF M PK100
<b>Manufacturing Date:</b>	October 2012
<b>Product shelf Life:</b>	September 2017

**Material:** Nitrile Butadiene Rubber, 100% latex Free

**QC Testing:** Representative production samples are collected and inspected in accordance with current applicable product specifications.

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)  
Meet or exceed EN 455-1&2  
ISO 2859 (sampling table)

**Accuracy:**

<b>Water tight test:</b>	Inspection level = G1; AQL1.5
<b>Visual inspection:</b>	Inspection level = G1; AQL1.5 (critical defects) AQL2.5 (major defects) AQL4.0 (minor defects)
<b>Dimensional:</b>	Inspection level = S2; AQL4.0
<b>Physical Property:</b>	Inspection level = S2; AQL4.0
<b>Powder Content:</b>	Inspection level = 5 pieces; Average $\leq 2$ mg/glove
<b>Moisture Content:</b>	Inspection level = 5 pieces; Average $\leq 1\%$
<b>Glove Count:</b>	Inspection level = 5 dispensers; Tolerance 100+/-2pieces

**Sterility:** N/A

**Non-Pyrogenic:** N/A

**Rnase/Dnase Free:** N/A

**Regulatory:** FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

**Signed:**

Ken Crossley  
Manager  
Quality Assurance

**Date of Issue:** October 22, 2012