

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance

Catalog Number: 7077-5N
Lot Number: 34112746
Product Description: PYREX® 5mL Disposable Serological Pipets, TD, Individually Wrapped, Sterile, Plugged

Component Materials:
Glass Type I Class B - Alumino-Borosilicate Glass, per ASTM E-438 (Standard Specification for Glasses in Laboratory Apparatus)
Plug: Polyester

Attributes: Capacity: 5mL
Sterility: Product Sterilized by ETO per ANSI/AAMI/ISO 11135-1994, ANSI/AAMI/ISO 11137-1994 and EN 550/552 requirements

This product is designed to comply with the requirements of ASTM E 714 (Specification for Disposable Glass Serological Pipets). This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following Characteristics:

Test/Procedure:	Result:
Visual Attributes	Pass
Dimensional Specification	Pass
Packaging	Pass
Capacity	Pass



Arthur D. Birdsall
Quality Assurance Manager

Rev No: 81372