



Quality Certificate

VWR* GLOVES

This document certifies that the below product meets the following criteria:

Product Number	82026-426
Lot Number	PT234R13E203
Description	VWR GLOVE NITRILE PF M PK100
Manufacturing Date:	June 2013
Product Shelf Life:	May 2018

Material: Nitrile Butadiene Rubber , 100% Latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current applicable product specification.

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)
Meet or exceed EN 455-1&2
ISO 2859 (sampling table)

Accuracy:

Water tight test:	Inspection level=G1; AQL 1.5
Visual inspection:	Inspection Level=G1;AQL 1.5 (criticaldefects) AQL 2.5 (major defects) AQL 4.0 (minor defects)
Dimensional:	Inspection Level= S2; AQL 4.0
Physical property:	Inspection Level=S2; AQL 4.0
Powder Content:	Inspection Level=5pieces ; Average \leq 2mg/glove
Moisture Content:	Inspection level=5pieces; Average \leq 1%
Glove Content:	Inspection level=5 dispensers; Tolerance 100+/-2 pieces

Sterility: N/A

Non-Pyrogenic: N/A

Rnase/Dnase Free: N/A

Regulatory: FDA 510K,ISO 9001,ISO 13485 under CMDCAS Quality System, GMP certificate

Signed:

Ken Crossley
Manager
Quality Assurance

Date: June 13, 2013