

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

# Certificate of Compliance

**Catalog Number:** 7065-1

**Lot Number:** 33412058

**Product Description:** PYREX® 1mL Reusable Measuring Pipets, 0.1mL Graduation Interval, TD, Color-Coded, Colored Markings

**Component Materials:** Type I Class B - Alumino-Borosilicate Glass, per ASTM E-438 (Standard Specification for Glasses in Laboratory Apparatus)

**Attributes:** Capacity: 1.0mL ±0.02mL

This product is designed to comply with the requirements of ASTM E 1293 (Specification for Glass Measuring Pipets). This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following Characteristics:

<b>Test/Procedure:</b>	<b>Result:</b>
Visual Attributes	Pass
Dimensional Specification	Pass
Packaging	Pass
Capacity	Pass



Arthur D. Birdsall  
Quality Assurance Manager

Rev No: 75426