



## Quality Certificate

---

### VWR\* SPECTROSIL SPECTROPHOTOMETER CELLS

---

This document certifies that the below product meets the following criteria:

<b>North American Catalog No:</b>	414004-062
<b>Lot Number:</b>	2012081921
<b>Description:</b>	VWR CELL QUARTZ FLAT 10MM PK2

**Material :** Quartz

**Quantity :** 6CS

INTERIOR		EXTERIOR			TRANSMISSION
Length	Width	Length	Width	Height	200nm>80%
10±0.05	10±0.2	12.5±0.1	12.5±0.1	45±0.5	

**Path Length :** 10mm

Signed:

Ken Crossley  
Manager  
Quality Assurance

Date of Issue: August 19, 2012