

Corning Incorporated
Life Sciences

Registered
ISO 9001:2008

Certificate of Compliance

Catalog Number: 1396-500
Lot Number: 17913001
Product Description: PYREX® Bottle, Media, Storage, Graduated, Square, GL-45 Thread, 500 mL,
with Orange PP Screw Cap and PP Pour Ring

Lot Number Designation:

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Component Materials:

Glass - Type I Class A - Low-Expansion Borosilicate Glass,
per ASTM E-438 (Standard Specification for Glasses in Laboratory Apparatus)
Cap - Polypropylene (PP)
Ring - Polypropylene (PP)

Attributes:

Capacity - 500 mL Approximate

TSE/BSE:

This product complies with the latest revision of EMEA/410/01, *Note for Guidance on Minimizing the Risk of Transmitting Animal Spongiform Encephalopathy Agents via Human and Veterinary Medicinal Product* by virtue that any component of the product which may include material of bovine origin has been processed per the specific conditions of section 6.4 of EMEA/410/01.

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

| Test/Procedure: | Results: |
|-------------------|----------|
| Visual Attributes | Pass |
| Dimensional | Pass |
| Packaging | Pass |

Revision Date:

17 JUN 2013



Arthur D. Birdsall
Quality Assurance Manager

Rev No: 1