



Certificate of Analysis

1.00181.0002 Oxidase for the detection of cytochrome oxidase in microorganisms
Bactident®

Batch HC374875

	Spec. Values	Batch Values
Cytochrome oxidase (Neisseria meningitidis ATCC 13090)	+	+
Cytochrome oxidase (Neisseria gonorrhoeae ATCC 9793)	+	+
Cytochrome oxidase (Pasteurella multocida ATCC 43137)	+	+
Cytochrome oxidase (Aeromonas hydrophila ATCC 7966 (WDCM 00063))	+	+
Cytochrome oxidase (Pseudomonas alcaligenes ATCC 14909)	+	+
Cytochrome oxidase (Escherichia coli ATCC 11775 (WDCM 00090))	-	-
Cytochrome oxidase (Staphylococcus aureus ATCC 25923 (WDCM 00034))	-	-
Cytochrome oxidase (Citrobacter freundii ATCC 8090)	-	-
Cytochrome oxidase (Serratia marcescens ATCC 14756)	-	-
Cytochrome oxidase (Proteus mirabilis ATCC 29906 (WDCM 00023))	-	-

Date of release (DD.MM.YYYY) 28.11.2013

Stefanie Fischer

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.