

BD Medical PO Box 749 Canaan CT 06018 US

CERTIFICATE OF COMPLIANCE

Page: 1 of 2

| Product Name | : SYRINGE 5ML LL SP125 |
|----------------|--|
| Catalog Number | : 309646 Manufacture Date : 2014/01/17 |
| Batch Number | : 4007481 |

REGULATORY COMPLIANCE AND QUALITY SYSTEM

BD Products comply with the regulatory requirements of the region in which these are sold and manufactured.

BD Products sold in the US comply with the current FDA Quality System Regulation 21CFR820. Medical devices are listed with FDA per 21CFR807. Manufacturing sites are registered with FDA per 21CFR807. The devices satisfy FDA pre-market notification requirements per 21CFR 807.

<u>STERILITY</u>

All products which are labeled as "sterile" and released for sale by BD are certified to be sterile as long as the package is unopened and undamaged.

This product is primarily sterilized via E-beam. Sterilization cycle development/validation is performed to 10-6 SAL in accordance with current ISO 11137 guidelines.

BIOCOMPATABILITY

This product has been evaluated in accordance with ISO 10993 "Biological Evaluation of Medical Devices", and complies with all relevant sections.

PYROGENICITY

All products which are labeled as non-pyrogenic and released for sale by BD have been tested per United States Pharmacopeia (USP) chapter 85 - Bacterial Endotoxins Test and meets limits as specified in chapter 161- Transfusion and Infusion Assemblies and Similar Medical Devices.

QUALITY CONTROL TESTING AND RELEASE

Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and signed off by qualified personnel for product



BD Medical PO Box 749 Canaan CT 06018 US

CERTIFICATE OF COMPLIANCE

Page: 2 of 2

| Product Name: SYRINGE 5ML LL SP125Catalog Number: 309646Manufacture Date : 2014/01/17Batch Number: 4007481 | |
|--|--|
| release. The released devices meet applicable BD product specifications. | |
| PRODUCT SPECIFIC SPECIFICATIONS | |
| BD MANUFACTURING SITE | |
| BD Manufacturing site: BD Medical, Route 7 and Grace Way, Canaan, CT 06018, USA | |
| Primary Sterilization site: BD Medical, Route 7 and Grace Way, Canaan, CT 06018, USA | |
| Legal Manufacturer: Becton, Dickinson and Company, 1 Becton Drive, Franklin Lakes, NJ 07417, USA | |
| File name 71002 CICH IC (NE Devicion Date 11/2E/2012 | |

File name Z1003_CLST_IS_CN6 Revision Date 11/25/2013

Rivers mary

Mary Rivers Quality Manager, BD Canaan