

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

# Certificate of Compliance

**Catalog Number:** 2116-50BO  
**Lot Number:** 30813042  
**Product Description:** PYREX® Buret, Body Only, Economy Grade, 50 mL

**Lot Number Designation:**

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Component Materials:**

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

**Attributes:**

Capacity - 50 mL  $\pm$ 0.10 mL at any mark

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

| <b>Test/Procedure:</b> | <b>Results:</b> |
|------------------------|-----------------|
| Visual Attributes      | Pass            |
| Dimensional            | Pass            |
| Packaging              | Pass            |
| Capacity               | Pass            |

**Revision Date:**

18 JUN 2013



Frigyes Szendro Kassa  
Quality Manager

Rev No: 1