

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance

Catalog Number: 431222

Lot Number: 24813065

Product Description: Corning® syringe filter, regenerated cellulose (RC) membrane, 0.20 µM pore size, 25 mm diameter, individually packaged

Component Materials:

- Housing - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.
- Membrane - Regenerated Cellulose (RC), 0.20 µM pore size, meets *USP, Class VI* requirements for plastic containers and closures.

Sterilization:

This lot has been irradiated and dosimetrically released based on the *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.
Sterility Assurance Level: SAL 10⁻⁵

Pyrogens:

The lot has been tested and has met the criteria established in the *USP Bacterial Endotoxin Test <85>*.
Results: less than 0.5 EU/mL

HPLC Certified:

The product has been tested to ensure there are no extraneous peaks in the UV range.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:


Test/Procedure:	Results:
Membrane Bubble Point Integrity	Pass
Housing Burst Pressure Test	Pass
Pressure Hold Test	Pass
Bacterial Challenge Test	Pass
Visual Attributes	Pass
Packaging	Pass

Expiration Date:

This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.

Revision Date:

9-11-12



Wendy Loitsch
Quality Manager

Rev No: 2