

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance

Catalog Number: 3022-1L
Lot Number: 34711051
Product Description: PYREX® Cylinder, Single Metric Scale, Graduated, TC, 1000 mL with LDPE Bumper Guard

Lot Number Designation:

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Component Materials:

Glass - Type I Class A - Low-Expansion Borosilicate Glass,
per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)
Guard - Low Density Polyethylene (LDPE)

Attributes:

Capacity - 1000 mL ±6.00 mL at any mark

This product is designed to comply with the requirements of ASTM E133 (Specification for Distillation Equipment), ASTM E1272 (Specification for Graduated Cylinders) and class B criteria per ASTM E694 (Specification for Volumetric Ware). It is calibrated in accordance with ASTM E542 (Specification for Calibration of Laboratory Volumetric Apparatus). This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure:

Visual Attributes
Dimensional
Packaging
Capacity

Results:

Pass
Pass
Pass
Pass

Revision Date:

03 JUL 2013



Frigyes Szendro Kassa
Quality Manager

Rev No: 1