

Corning Incorporated  
Life Sciences

Registered  
ISO 9001:2008

# Certificate of Compliance

**Catalog Number:** 7077-5N  
**Lot Number:** 06112746  
**Product Description:** PYREX® Pipet, Disposable Glass, Serological, 5 mL in 1/10 mL Individual Wrap, Sterile, Plugged

**Lot Number Designation:**

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Component Materials:**

Glass - Type I Class B - Alumino-Borosilicate Glass,  
per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)  
Plug - Polyester (13 grain)

**Attributes:**

Capacity - 5.0 mL ± 0.15 mL at any mark (3% of full capacity)

**Sterilization:**

Product sterilized by Ethylene Oxide (ETO) per ANSI/AAMI/ISO 11135-1994, ANSI/AAMI/ISO 11137-1994 and EN 550/552 requirements.

**TSE/BSE:**

Materials of animal origin are not used in the manufacture of this product.

This product is designed to comply with the requirements of ASTM E 714 (Specification for Disposable Glass Serological Pipets). This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

**Test/Procedure:**

Visual Attributes  
Dimensional  
Packaging  
Capacity

**Results:**

Pass  
Pass  
Pass  
Pass

**Revision Date:**

9 OCT 2013



Arthur D. Birdsall  
Quality Assurance Manager

Rev No: 2