

Corning Incorporated
Life Sciences

Registered
ISO 9001:2008

Certificate of Compliance

Catalog Number: 5100-500
Lot Number: 01513002
Product Description: PYREX® Flask, Erlenmeyer, Heavy Duty Rim, Wide Mouth, Graduated, 500 mL
Tooled for Rubber Stopper #10

Lot Number Designation:

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Component Materials:

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

Attributes:

Capacity - 500 mL ±25 mL at any mark (5% of full capacity)

TSE/BSE:

Materials of animal origin are not used in the manufacture of this product.

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Visual Attributes	Pass
Dimensional	Pass
Packaging	Pass

Revision Date:
11 MAR 2013



Arthur D. Birdsall
Quality Assurance Manager

Rev No: 1