



Quality Certificate

VWR * NITRILE EXAMINATION GLOVE

This document certifies that the below product meets the following criteria:

North American Catalog No:	82026-426
Lot No:	PT234R14C002
Description:	VWR GLOVE NITRILE PF M PK100
Manufacturing Date :	March 2014
Product Shelf Life :	February 2019

Material: Nitril BUTADIENE Rubber, 100% Latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

Accuracy:	Water tight test:	Inspection level = G1;	AQL 1.5
	Visual inspection:	Inspection level = G1;	AQL 1.5 (critical defects) AQL 2.5 (major defects) AQL 4.0 (minor defects)
	Dimensional:	Inspection level = S2;	AQL 4.0
	Physical Property:	Inspection level = S2;	AQL 4.0
	Powder Content:	Inspection level = 5 pieces;	Average ≤ 2 mg/glove
	Moisture Content:	Inspection level = 5 pieces;	Average ≤ 1%
	Glove count:	Inspection level = 5 dispensers;	Tolerance 100+/-2 pieces

Sterility N/A

Non- Pyrogenic: N/A

Rnase/ Dnase Free: N/A

Regulatory: FDA 510K,ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

Signed:



Ken Crossley
Manager
Quality Assurance

Date: March 28, 2013