

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Certificate of Compliance

Catalog Number: 4985-250
Lot Number: 13914052
Product Description: PYREX® Flask, Erlenmeyer, Graduated, 250 mL
with G.P.I. 430-38 Phenolic Cap and Rubber Liner

Lot Number Designation:

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Component Materials:

Glass - Type I Class A - Low-Expansion Borosilicate Glass,
per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)
Cap - Phenolic
Liner - Rubber

Attributes:

Capacity - 250 mL ±12 mL at any mark (5% of full capacity)

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure:

Visual Attributes
Dimensional
Packaging

Results:

Pass
Pass
Pass

Revision Date:

19 JUL 2013



Racelis Acosta Reynoso
Plant Quality Manager

Rev No: 1