



## Quality Certificate

### VWR\* GLOVES

This document certifies that the below product meets the following criteria:

|                                   |                              |
|-----------------------------------|------------------------------|
| <b>North American Catalog No:</b> | 82026-424                    |
| <b>Lot Number:</b>                | PT234R13C001                 |
| <b>Description:</b>               | VWR GLOVE NITRILE PF S PK100 |
| <b>Manufacturing Date:</b>        | March 2013                   |
| <b>Product shelf Life:</b>        | February 2018                |

**Material:** Nitrile Butadiene Rubber, 100% latex Free

**QC Testing:** Representative production samples are collected and inspected in accordance with current applicable product specifications.

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)  
Meet or exceed EN 455-1&2  
ISO 2859 (sampling table)

**Accuracy:**

|                           |  |
|---------------------------|--|
| <b>Water tight test:</b>  | Inspection level = G1; AQL1.5  |
| <b>Visual inspection:</b> | Inspection level = G1; AQL1.5 (critical defects)<br>AQL2.5 (major defects)<br>AQL4.0 (minor defects) |
| <b>Dimensional:</b>       | Inspection level = S2; AQL4.0  |
| <b>Physical Property:</b> | Inspection level = S2; AQL4.0  |
| <b>Powder Content:</b>    | Inspection level = 5pieces; Average $\leq 2$ mg/glove  |
| <b>Moisture Content:</b>  | Inspection level = 5pieces; Average $\leq 1\%$   |
| <b>Glove Count:</b>       | Inspection level = 5 dispensers; Tolerance 100+/-2pieces   |

**Sterility:** N/A

**Non-Pyrogenic:** N/A

**Rnase/Dnase Free:** N/A

**Regulatory:** FDA510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

Signed:

Ken Crossley  
Manager  
Quality Assurance

Date of Issue: March 31, 2013