



NIPRO MEDICAL INDUSTRIES LTD.

2-19-64, Matsubara, Tatebayashi-shi, Gunma, 374-8518 Japan
TEL:0276-72-5117 FAX:0276-74-6723

TO : Becton Dickinson Vacutainer System
Attention : BD PAS Quality Assurance Department

Date : Feb 5,2014

Certificate of Conformance

Purchase Order No. : 4503127869
Reorder No. : 367296
Lot No. : 4A2581
Quantity : 132000 pcs
Date of Manufacture : Jan 28,2014
Expiry : 2017-01
Specification : VS59838
Revision : 19

This is to certify that the item identified herein has been tested in accordance with the specification of VS59838 and meets its requirements. It is further certified that no material/process substitutions were made for any components specified in VS59838.

Test	Sample	AQL	Judgement
Biological Indicator	22 BI/Chamber	0	Passed
Bacterial Endotoxins	10 Set	< 0.5EU/mL	Passed

Method

Biological Indicator

U.S. Pharmacopeia 36:2013.

<71>Sterility Test

Bacterial Endotoxins

U.S. Pharmacopeia 36:2013.

<85>Bacterial Endotoxins Test

<161>Transfusion and Infusion Assemblies

Monitoring	Tolerance	Judgement
Process parameters	Within limit	Passed

Method :ISO11135-1:2007.

Authorized Signature :

Minoru Oishi
Quality Control Manager
Nipro Medical Industries Ltd.