



Quality Certificate

VWR * NITRILE EXAMINATION GLOVE

This document certifies that the below product meets the following criteria:

| | |
|-----------------------------------|------------------------------|
| North American Catalog No: | 82026-428 |
| Lot No: | PT234R14K202 |
| Description: | VWR GLOVE NITRILE PF L PK100 |
| Manufacturing Date: | August 2014 |
| Product Shelf Life : | July 2019 |

Material: Nitrile BUTADIENE Rubber, 100% Latex Free

QC Testing: Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

| | | | |
|------------------|--------------------|----------------------------------|--|
| Accuracy: | Water tight test: | Inspection level = G1; | AQL 1.5 |
| | Visual inspection: | Inspection level = G1; | AQL 1.5 (critical defects) AQL 2.5 (major defects) AQL 4.0 (minor defects) |
| | Dimensional: | Inspection level = S2; | AQL 4.0 |
| | Physical Property: | Inspection level = S2; | AQL 4.0 |
| | Powder Content: | Inspection level = 5 pieces; | Average ≤ 2 mg/glove |
| | Moisture Content: | Inspection level = 5 pieces; | Average ≤ 1% |
| | Glove count: | Inspection level = 5 dispensers; | Tolerance 100+/-2 pieces |

Sterility N/A

Non- Pyrogenic: N/A

Rnase/ Dnase Free: N/A

BSE/ TSE Statement: The Manufacturer declares that this product is free from BSE (Bovine,Spongiform Encephalopathy) and TSE (Transmissible Spongiform Encephalopathy),Since this product contains or was produced from no animal by product or animal derived ingredients.

Regulatory: FDA 510K,ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

Signed:



Ken Crossley
Manager
Quality Assurance

Date: August 26, 2014