

Corning Incorporated  
Life Sciences

Registered  
ISO 9001:2008

# Certificate of Compliance

**Catalog Number:** 1396-500  
**Lot Number:** 09314001  
**Product Description:** PYREX® Bottle, Media, Storage, Graduated, Square, GL-45 Thread, 500 mL,  
with Orange PP Screw Cap and PP Pour Ring

**Lot Number Designation:**

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Component Materials:**

Glass - Type I Class A - Low-Expansion Borosilicate Glass,  
per ASTM E-438 (Standard Specification for Glasses in Laboratory Apparatus)  
Cap - Polypropylene (PP)  
Ring - Polypropylene (PP)

**Attributes:**

Capacity - 500 mL Approximate

**TSE/BSE:**

This product complies with the latest revision of EMEA/410/01, *Note for Guidance on Minimizing the Risk of Transmitting Animal Spongiform Encephalopathy Agents via Human and Veterinary Medicinal Product* by virtue that any component of the product which may include material of bovine origin has been processed per the specific conditions of section 6.4 of EMEA/410/01.

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

Test/Procedure:	Results:
Visual Attributes	Pass
Dimensional	Pass
Packaging	Pass

**Revision Date:**  
17 JUN 2013



Arthur D. Birdsall  
Quality Assurance Manager

Rev No: 1