

Greiner Bio-One North America, Inc. 4238 Capital Drive Monroe NC, 28110, USA

Certificate of Compliance

Product description:	VACUETTE ®	Blood Collection	Tube
----------------------	-------------------	------------------	------

3 ml Z No Additive

Greiner product No.:

454241

Lot No.:

B140713S

Expiry date:

January - 2016

Sterility:

Sterile

1 Visual checks passed

2 Physical tests passed

We hereby certify that the tubes were tested in accordance with the control procedures stated in the following international standards and meet the required specifications:

EN 14820

"Single-use containers for venous blood specimen collection"

ISO 11137

"Sterilization of health care products – Requirements for validation and

routine control - Radiation sterilization"

The product has been manufactured under controlled conditions in accordance with ISO 9001 and EN ISO 13485.

Nov. II, 2019 Date B. QM S

on behalf of

QA / QC Manager