

Greiner Bio-One North America, Inc.  
4238 Capital Drive  
Monroe NC, 28110, USA



**Certificate of Compliance**

Product description: VACUETTE® Blood Collection Tube  
7 ml Z Serum Clot Activator with Gel

Greiner product No.: 455099  
Lot No.: B041404  
Expiry date: October – 2015

Sterility: Sterile

1	Visual checks	<input checked="" type="checkbox"/>	passed
2	Physical tests	<input checked="" type="checkbox"/>	passed
3	Chemical tests	<input checked="" type="checkbox"/>	passed

We hereby certify that the tubes were tested in accordance with the control procedures stated in the following international standards and meet the required specifications:

EN 14820 "Single-use containers for venous blood specimen collection"  
ISO 11137 "Sterilization of health care products – Requirements for validation and routine control – Radiation sterilization"

The product has been manufactured under controlled conditions in accordance with ISO 9001 and EN ISO 13485.

Feb 24, 2015  
Date



J. Guthrie on behalf of  
Carolyn Guthrie  
QA / QC Manager