Greiner Bio-One North America, Inc. 4238 Capital Drive Monroe NC, 28110, USA



Certificate of Compliance

Product description:	VACUETTE®	Blood	Collection	Tube
----------------------	------------------	-------	------------	------

2ml Sodium Heparin Tube

Greiner product No.:

454302

Lot No.:

B021401

Expiry date:

August - 2015

Sterility:

Sterile

	1	Visual checks		passed
	2	Physical tests	\boxtimes	passed
;	3	Chemical tests	\boxtimes	passed

We hereby certify that the tubes were tested in accordance with the control procedures stated in the following international standards and meet the required specifications:

ISO 6710

"Single-use containers for venous blood specimen collection"

ISO 11137

"Sterilization of health care products - Requirements for validation and

routine control - Radiation sterilization"

The product has been manufactured under controlled conditions in accordance with ISO 9001 and EN ISO 13485.

Feb 24, 2017
Date



on behalf of

Carolyn Guthrie Quality Manager