



Becton Dickinson Caribe LTD.
 PO Box 372860
 Cayey PR 00737-2860 PR

Certificate of Analysis

Product Name : DISC CHLORAMPHENICOL C-30 10 EA
Catalog Number : 231274 **Manufacture Date** : 2011/04/25
Batch Number : 1105156
Expiration Date : 2015/04/30

Characteristic	Unit	Value	LowLimit	HighLimit
Micro Assay Range:		31.37		
Micro Assay Range:		30.94		
Micro Assay Range:		29.63		
Micro Assay Range:		30.68		
Micro Assay Potenc		31.00		
Uniformity		1.4	-	2.5
Uniformity		1.0	-	2.5
Uniformity		1.0	-	2.5
Uniformity		0.8	-	2.5

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostic Systems (BDDS) is an ISO 13485:2003 and ISO 9001:2008 Registered facility. BDDS products are manufactured in facilities registered with the United States Food and Drug Administration (FDA), and are regulated by the FDA's Quality System Regulations (QSRs). This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release.

John Gerlich
 Vice President,
 Quality Management and
 Regulatory Compliance
 Signature Date: 2011/05/23