



Quality Certificate

VWR® NITRILE EXAMINATION GLOVE

This document certifies that the below product meets the following criteria:

North American Catalog No :	82026-426
Lot No :	PT234R14N103
Description :	VWR GLOVE NITRILE PF M PK100
Date of Manufacturing :	October 2014
Product Shelf Life :	September 2019
Country of Origin :	Indonesia
Quality System Compliance :	FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

QC Testing: Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

Accuracy:

Water tight test:	Inspection level = G1;	AQL 1.5
Visual inspection:	Inspection level = G1;	AQL 1.5 (critical defects)
		AQL 2.5 (major defects)
		AQL 4.0 (minor defects)
Dimensional:	Inspection level = S2;	AQL 4.0
Physical Property:	Inspection level = S2;	AQL 4.0
Powder Content:	Inspection level = 5 pieces;	Average ≤ 2 mg/glove
Moisture Content:	Inspection level = 5 pieces;	Average ≤ 1%
Glove count:	Inspection level = 5 dispensers;	Tolerance 100+/-2 pieces

Product Specifications:

Material: Nitril BUTADIENE Rubber, 100% Latex Free

Sterility: N/A

Non- Pyrogenic: N/A

RNase/ DNase Free: N/A

Cytotoxicity: N/A

BSE/ TSE Statement: N/A

Signed :



Ken Crossley
Manager
Quality Assurance

Date : October 24, 2014