

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

# Certificate of Compliance

**Catalog Number:** 4985-250  
**Lot Number:** 19014121  
**Product Description:** PYREX® Flask, Erlenmeyer, Graduated, 250 mL  
with G.P.I. 430-38 Phenolic Cap and Rubber Liner

**Lot Number Designation:**

This production lot has been uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Component Materials:**

Glass - Type I Class A - Low-Expansion Borosilicate Glass,  
per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)  
Cap - Phenolic  
Liner - Rubber

**Attributes:**

Capacity - 250 mL ±12 mL at any mark (5% of full capacity)

This product has been sampled in accordance with Standard Operating Procedure and has been released by Quality Assurance for the following characteristics:

**Test/Procedure:**

Visual Attributes  
Dimensional  
Packaging

**Results:**

Pass  
Pass  
Pass

**Revision Date:**

19 JUL 2013



Racelis Acosta Reynoso  
Plant Quality Manager

Rev No: 1