



## Quality Certificate

### VWR® NITRILE EXAMINATION GLOVE

This document certifies that the below product meets the following criteria:

<b>North American Catalog No :</b>	82026-428
<b>Lot No :</b>	PT234R14S007
<b>Description :</b>	VWR GLOVE NITRILE PF L PK100
<b>Date of Manufacturing :</b>	December 2014
<b>Product Shelf Life :</b>	November 2019
<b>Country of Origin :</b>	Indonesia
<b>Quality System Compliance :</b>	FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

**QC Testing:** Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

**Accuracy:**

Water tight test:	Inspection level = G1;	AQL 1.5
Visual inspection:	Inspection level = G1;	AQL 1.5 (critical defects)
		AQL 2.5 (major defects)
		AQL 4.0 (minor defects)
Dimensional:	Inspection level = S2;	AQL 4.0
Physical Property:	Inspection level = S2;	AQL 4.0
Powder Content:	Inspection level = 5 pieces;	Average ≤ 2 mg/glove
Moisture Content:	Inspection level = 5 pieces;	Average ≤ 1%
Glove count:	Inspection level = 5 dispensers;	Tolerance 100+/-2 pieces

#### Product Specifications:

**Material:** Nitril BUTADIENE Rubber, 100% Latex Free

**Sterility:** N/A

**Non- Pyrogenic:** N/A

**RNase/ DNase Free:** N/A

**Cytotoxicity:** N/A

**BSE/ TSE Statement:** N/A

**Signed :**



Ken Crossley  
Manager  
Quality Assurance

**Date :** 26 Dec 2014