



Quality Certificate

VWR® SPECTROSIL SPECTROPHOTOMETER CELLS

This document certifies that the below product meets the following criteria:

North American Catalog No:	414004-062
Lot Number:	2013031621
Description:	VWR CELL QUARTZ FLAT 10MM PK2

Material : Quartz

INTERIOR		EXTERIOR			TRANSMISSION (at 200nm)
Path Length	Path Width	Length	Width	Height	>80%
10±0.05mm	10±0.2mm	12.5±0.1mm	12.5±0.1mm	45±0.5mm	

Signed:

Ken Crossley
Manager
Quality Assurance

Date of Issue: March 16, 2013