



## Quality Certificate

---

### VWR® PETRI DISH

---

Product No.	Description	Mfg. Lot No.	Sterility Lot No.
470210-576	VWR PETRI DISH 150X15MM CS100	I2525152	G0926152

As per the manufacturer, the above listed product has been inspected and released in accordance with the plant standard production methods and process quality assurance specifications. The resins used in the processing of this product has been approved and released.

The product listed above has been processed for sterility by Electron-Beam Radiation in order to achieve Sterility Assurance Level (SAL) of  $10^{-3}$  as a minimum, in accordance with the below guidelines and Standards:

- Title 21 CFR Part 820 US FDA Good Manufacturing Practice – Quality Systems Regulations (June 1997)
- ISO 13485:2003 Quality Management System for Medical Devices
- ISO 11137:2006 Sterilization of Healthcare Products

**Signed:**

Ken Crossley  
Manager  
Quality Assurance

**Date:** November 06, 2015