


Date of Certificate: 09/02/2015
Product Number: 5000-0020
Product Lot Number: 1154779
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1560579	08/13/2015	312	CS	20.3	25.4
1567590	08/18/2015	312	CS	19.7	26.4
1563966	08/18/2015	156	CS	20.5	25.3
1562227	08/19/2015	156	CS	20.2	25.1
1570165	08/21/2015	156	CS	20.0	26.2
1575831	08/31/2015	156	CS	19.9	25.4



Barry Srolis
Sr. Quality Manager