

Certificate of Conformity

Product: Nunc Flasks, Vent/Close Caps, 25 cm2

Date: 2015-06-10

Page: 1

Item number: 163371

Lot number: 143923

Date of manufacture: May 2015

Date of expiry: May 2020

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies
Cloning efficiency with V-79 cell line	According to specification	Complies
USP <85> Bacterial endotoxins test	<0.5 endotoxins unit/ml	Complies
USP <88> "Biological Reactivity test In Vivo, USP Class VI Test"	USP current edition	Complies**
Cell Culture Test L 929 Human Lung Primary chicken embryo	According to specification	Complies

**Only performed once

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation", with SAL 10E-6.

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2008 "Quality management system - Requirements" and ISO 13485:2003 "Medical devices - Quality management systems - System requirements for regulatory purposes".

Roskilde Site



Anne-Mette Reimann
QA Manager

Nunc A/S
Kamstrupvej 90
P.O.Box 280
DK-4000 Roskilde
Denmark

+45 4631 2000 phone
+45 4631 2042 fax
Info.nunc@thermofisher.com

www.thermofisher.com

A/S 24482 VAT/SE NO. DK 39582910