

CERTIFICATE OF STERILITY

VIA ASEPTIC PROCESSING

STERILE

A

Date: 2016-03-31

Product no : D210-16

Lot no : B25751362

Sterility expiration date: 09/2020

All specified product lots have been manufactured using an aseptic process and tested according to USP <71>. Each level has been validated to secure that all parts meet a sterility assurance level of a minimum of SAL 10^{-3} .

Annette Roy

QA coordinator / coordonatrice AQ