



Quality Certificate

VWR* PETRI DISH

Product No.	Description	Lot No.	Sterility Lot No.
25384-302	VWR PETRI DISH 100X15MM CS500	G2626131	G0726132

This is to certify that the above mentioned product has been inspected and released in accordance with the plant standard production methods and process quality assurance specifications. All resins used in the processing of these products have been approved and released.

The products listed above have been processed for sterility by Electron-Beam Radiation in order to achieve Sterility Assurance Level (SAL) of 10^{-3} as a minimum, in accordance with the below guidelines and Standards:

- Title 21 CFR Part 820 US FDA Good Manufacturing Practice – Quality Systems Regulations (June 1997)
- ISO 13485:2003 Quality Management System for Medical Devices
- ISO 11137:2006 Sterilization of Healthcare Products

Signed:

A handwritten signature in blue ink, appearing to read "Ken Crossley".

Ken Crossley
Manager
Quality Assurance

Date of Issue: Sep 04, 2013