



Quality Certificate

VWR® PETRI DISH

| Product No. | Description | Lot No. | Sterility Lot No. |
|-------------|----------------------------------|----------|-------------------|
| 25384-302 | VWR PETRI DISH 100X15MM CS500 | G0325132 | C232713 |

This is to certify that the above mentioned product has been inspected and released in accordance with the plant standard production methods and process quality assurance specifications. All resins used in the processing of these products have been approved and released.

The products listed above have been processed for sterility by Electron-Beam Radiation in order to achieve Sterility Assurance Level (SAL) of 10^{-3} as a minimum, in accordance with the below guidelines and Standards:

- Title 21 CFR Part 820 US FDA Good Manufacturing Practice – Quality Systems Regulations (June 1997)
- ISO 13485:2003 Quality Management System for Medical Devices
- ISO 11137:2006 Sterilization of Healthcare Products

Signed:

A handwritten signature in blue ink, appearing to read 'Ken Crossley', written over a light blue horizontal line.

Ken Crossley
Manager
Quality Assurance

Date of Issue: March 27, 2013