

Greiner Bio-One North America, Inc.
4238 Capital Drive
Monroe NC, 28110, USA



Certificate of Compliance

Product description: Safety Blood Collect with Luer Adapter
21 x 12"

Greiner product No.: 450095
Lot No.: 08H30
Quantity: 173,000 pieces

Sterility: Sterile

1	Visual checks	<input checked="" type="checkbox"/>	passed
2	Physical tests	<input checked="" type="checkbox"/>	passed
3	Chemical tests	<input checked="" type="checkbox"/>	passed


We hereby certify that the needles were tested in accordance with the control procedures stated in the following standards and meet the required specifications:

ISO 11135 "Medical Devices – Validation and Routine Control of Ethylene Oxide Sterilization"
EN 550 "Sterilization of Medical Devices – Validation and Routine Control of Ethylene Oxide Sterilization"

The product has been manufactured under controlled conditions in accordance with ISO 9001 and EN ISO 13485.

We herewith confirm that these needles are siliconed.

4-2-09
Date


on behalf of
Manfred Abel
Quality Manager