

Greiner Bio-One North America, Inc.  
4238 Capital Drive  
Monroe NC, 28110, USA



**Certificate of Compliance**

Product description: VACUETTE® Blood Collection Tube  
4 ml K3E K<sub>3</sub>EDTA

Greiner product No.: 454012  
Lot No.: B14053BX  
Expiry date: November - 2015

Sterility: Sterile

1	Visual checks	<input checked="" type="checkbox"/>	passed
2	Physical tests	<input checked="" type="checkbox"/>	passed
3	Chemical tests	<input checked="" type="checkbox"/>	passed

We hereby certify that the tubes were tested in accordance with the control procedures stated in the following international standards and meet the required specifications:

EN 14820 "Single-use containers for venous blood specimen collection"  
ISO 11137 "Sterilization of health care products – Requirements for validation and routine control – Radiation sterilization"

The product has been manufactured under controlled conditions in accordance with ISO 9001 and EN ISO 13485.

March 31, 2015  
Date



[Signature] on behalf of  
Carolyn Guthrie  
QA / QC Manager