

Certificate of Analysis

Description: HBSS, 1X
(Hank's Balanced Salt Solution)
 with calcium & magnesium

Catalog #: 21-020-CV
Lot #: 21020145
Expiration Date: 09/17
Country of Origin/Final Processing: USA

Storage: 2-30°C.

Notes:

Mediatech, Inc. products are prepared by a validated aseptic sterile filtration process to ensure that all products comply with the industry's Sterility Assurance Level (SAL) of at least 10^{-3} (i.e. demonstrating that a product produced in this manner has no more than 1 random contaminant per 1000 units).

This product was tested in accordance with currently approved Mediatech, Inc. specifications and procedures. Testing procedures are maintained in compliance with the current versions of the USP and/or EP, where applicable.

For IVD Use. Not for Human Therapeutic Use. Utilization of this product apart from the labeled intended use may be a violation of local and/or Federal Law.

TEST PARAMETERS	SPECIFICATION	RESULT
pH	7.1 – 7.4	7.2
Osmolality	275 ± 15 mOsm/Kg H ₂ O	282 mOsm/Kg H ₂ O
Endotoxin	≤ 0.25 EU/mL @ FW	<0.0010 EU/mL @ FW
Mycoplasma	Tested Negative	Negative
Sterility	Pass	Pass

For detailed information on Mediatech, Inc. testing procedures, please contact our Technical Services department at (800) (235-5476) or via email at MTTechServ@corning.com.

Written By/Date:

[Signature] 29 Oct 2014
 Quality Systems Department

Verified By/ Date:

[Signature] / 31 Oct 2014