

Corning Incorporated - Life Sciences  
Antigua Carretera A Roma KM 3.6  
San Nicolas De Los Garza NL 66480 MEX  
www.corning.com/lifesciences  
Refer to website for regional contact information.

Page: 1 / 1

<b>Product Name</b> : 7065 PIPET 1-1/10	<b>Manufacture Date</b> : 2015-07-08
<b>Catalog Number</b> : 7065-1	
<b>Lot ID</b> : 18915031	
<b>Expiration Date</b> : Unconfirmed Validity	

**Quality Management System** - Complies with the current version of the ISO 9001 Standard.

**Glass Type** - Alumino-Borosilicate, Type I, Class B, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

**Volumetric Tolerance** - Volumetric Accuracy at any graduation +/- 0.02 mL.

**Product Specification Standards** - Product made in accordance with:

- ASTM E542 - Specification for Calibration of Laboratory Volumetric Apparatus
- ASTM E1293 - Standard Specification for Glass Measuring Pipets

**BSE/TSE** - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

- Visual Inspection - Pass
- Packaging Inspection - Pass
- Dimensional Test - Pass
- Capacity Verification - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Racelis Acosta Reynoso  
Quality Manager