



Quality Certificate

VWR® GRADUATED CYLINDER

As per the manufacturer, the below mentioned product meets the following criteria:

North American Catalog No:	65000-012
Lot Number:	J267347
Description:	VWR CYLINDER POLYPRO 1000ML
Date of Manufacture:	11/22-11/24/2015
Material:	Polypropylene, Clarified, FDA 21 CFR177.1520
Country of Origin:	United States of America

This is to verify that all items noted below are in conformance with the drawings, specifications, and other applicable documentation.

BSE/TSE Statement: BSE/TSE protein free

Other Compliances:

- Graduated cylinders have a total contained, total delivered scale on one side.
- Large spout facilitates pouring and hexagon base provides stability.
- Non wetting surface eliminates concave meniscus and gives more accurate readings.
- Cylinders are chemical resistant and steam autoclavable at 121° C.
- +/- 2ml / 5% accuracy.

Signed:

Ken Crossley
Manager
Quality Assurance

Date: May 19, 2016