

Corning Incorporated - Life Sciences  
Antigua Carretera A Roma KM 3.6  
San Nicolas De Los Garza NL 66480 MEX  
www.corning.com/lifesciences  
Refer to website for regional contact information.

Page: 1 / 1

<b>Product Name</b> : 4985 FLASK 250ML	<b>Manufacture Date</b> : 2016-05-21
<b>Catalog Number</b> : 4985-250	
<b>Lot ID</b> : 14216020	
<b>Expiration Date</b> : Unconfirmed Validity	

**Quality Management System** - Complies with the current version of the ISO 9001 Standard.

**Glass Type** - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

**Volumetric Tolerance** - Volumetric Accuracy at any graduation +/- 5 % of full capacity.

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

- Visual Inspection - Pass
- Packaging Inspection - Pass
- Dimensional Test - Pass
- Capacity Verification - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Racelis Acosta Reynoso  
Quality Manager