



Quality Certificate

VWR® SINGLE SCALE GRADUATED CYLINDERS, POLYPROPYLENE

As per the manufacturer, the below product meets the following criteria:

North American Catalog No:	65000-012
Lot Number:	J281845
Description:	VWR CYLINDER POLYPRO 1000ML
Date of Manufacture:	July 12, 2016
Country of Origin:	United States of America

Quality Conformance

This is to verify that all items noted below are in conformance with the Purchase Order Contract, drawings, specifications, and other applicable documentation.

Product Specifications

Material : Polipropylene, Clarified, FDA 21 CFR177.1520

BSE/TSE Statement: BSE/TSE protein free

Other Compliances:

- Graduated cylinder have a total contained, total delivered scale on one side.
- Large spout facilities pouring and hexagon base provides stability.
- Non wetting surface eliminates concave meniscus and gives more accurate readings.
- Cylinders are chemical resistan and steam autoclavable at 121°C +/- 2 ml / 5% accuracy.

Signed:



Ken Crossley
Manager
Quality Assurance

Date: August 03, 2016