



Quality Certificate

VWR® NITRILE EXAMINATION GLOVE

As per the manufacturer, the below product meets the following criteria:

North American Catalog No :	82026-428
Lot No :	PT16A106
Description :	VWR GLOVE NITRILE PF L PK100
Date of Manufacturing :	Jan 2016
Product Shelf Life :	Dec 2020
Country of Origin :	Indonesia
Quality System Compliance :	FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

QC Testing: Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

Accuracy:	Water tight test:	Inspection level = G1;	AQL 1.5
	Visual inspection:	Inspection level = G1;	AQL 1.5 (critical defects) AQL 2.5 (major defects) AQL 4.0 (minor defects)
	Dimensional:	Inspection level = S2;	AQL 4.0
	Physical Property:	Inspection level = S2;	AQL 4.0
	Powder Content:	Inspection level = 5 pieces;	Average ≤ 2 mg/glove
	Moisture Content:	Inspection level = 5 pieces;	Average ≤ 1%
	Glove count:	Inspection level = 5 dispensers;	Tolerance 100+/-2 pieces

Product Specifications:

Material:	Nitrile Butadiene Rubber, Not made with natural rubber latex.
Sterility	N/A
Non- Pyrogenic:	N/A
RNase/ DNase Free:	N/A
Cytotoxicity:	N/A
BSE/ TSE Statement:	N/A

Signed :



Ken Crossley
Manager
Quality Assurance

Date : Jan 16, 2016