

## Certificate of Quality

**REF**

421180

**LOT**

E16053CW

Product description: SWAB TUBE, PS, 16/152 MM, COTTON SWAB, STERILE,  
SINGLE PACKED

Following inspections were carried out and passed:

|              |                                     |        |
|--------------|-------------------------------------|--------|
| raw material | <input checked="" type="checkbox"/> | passed |
| visual       | <input checked="" type="checkbox"/> | passed |
| physical     | <input checked="" type="checkbox"/> | passed |

We hereby certify that our products were manufactured and tested in accordance with the following international standards and meet the required specifications:

ISO 11137 "Sterilization of health care products - Radiation"

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.



2021-05-03

**STERILE**

SAL  $10^{-3}$

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.



Certificate creation date: 2016-11-11

[www.gbo.com](http://www.gbo.com)