

Date of Certificate: 11/10/2016
Product Number: 5000-0020
Product Lot Number: 1187784
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1835404	11/08/2016	156	CS	20.0	25.4
1836602	11/08/2016	156	CS	19.7	25.6



Barry Srolis
Sr. Quality Manager