


---

Date of Certificate: 01/06/2017  
Product Number: 5000-0020  
Product Lot Number: 1191463  
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1871130	01/05/2017	132	CS	20.0	25.5



---

**Barry Srolis**  
Sr. Quality Manager