

Date of Certificate: 02/15/2017
Product Number: 5000-0020
Product Lot Number: 1193628
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1886842	01/31/2017	25	CS	19.8	25.7
1888022	01/31/2017	156	CS	20.3	25.0
1891690	02/14/2017	131	CS	19.9	26.2



Barry Srolis
Sr. Quality Manager