

Date of Certificate: 02/02/2017
Product Number: 5000-0020
Product Lot Number: 1191464
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
1874859	01/23/2017	24	CS	20.2	25.2
1883463	01/23/2017	156	CS	20.4	25.6
1883477	01/24/2017	29	CS	20.1	25.8
1884989	02/01/2017	102	CS	21.4	26.3



Barry Srolis
Sr. Quality Manager