



## Quality Certificate

### VWR® POWDER-FREE NITRILE EXAMINATION GLOVES

As per the manufacturer, the below product meets the following criteria:

<b>North American Catalog No :</b>	82026-426
<b>Lot No :</b>	PT15K004
<b>Description :</b>	VWR GLOVE NITRILE PF M PK100
<b>Date of Manufacturing :</b>	August 2015
<b>Product Shelf Life :</b>	July 2020
<b>Country of Origin :</b>	Indonesia
<b>Quality System Compliance :</b>	FDA 510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

**QC Testing:** Representative production samples are collected and inspected in accordance with current applicable Product specification:

- Meet or exceed ASTM D 6319-00a (Water test & Dimension)
- Meet or exceed EN 455 -1&2
- ISO 2859 (Sampling table)

**Accuracy:**

Water tight test:	Inspection level = G1;	AQL 1.5
Visual inspection:	Inspection level = G1;	AQL 1.5 (critical defects)
		AQL 2.5 (major defects)
		AQL 4.0 (minor defects)
Dimensional:	Inspection level = S2;	AQL 4.0
Physical Property:	Inspection level = S2;	AQL 4.0
Powder Content:	Inspection level = 5 pieces;	Average ≤ 2 mg/glove
Moisture Content:	Inspection level = 5 pieces;	Average ≤ 1%
Glove count:	Inspection level = 5 dispensers;	Tolerance 100+/-2 pieces

#### Product Specifications:

<b>Material:</b>	Nitrile Butadiene Rubber, Not made with natural rubber latex.
<b>Sterility</b>	N/A
<b>Non- Pyrogenic:</b>	N/A
<b>RNase/ DNase Free:</b>	N/A
<b>Cytotoxicity:</b>	N/A
<b>BSE/ TSE Statement:</b>	N/A

#### Signed:



Linda K. Shine  
Manager - Supplier Qualification  
Global Compliance  
VWR International LLC

**Date :** August 13, 2015